

PRODUCT SPECIFICATIONS

Date: 2015-05-12

| | | |
|-------------------------------|---|--|
| PRODUCT NAME | : | Anti-h Thyroglobulin 2805 SPTN-5 |
| PRODUCT SPECIFICITY | : | Human Thyroglobulin |
| PRODUCT CODE | : | 100334 |
| PRODUCT BUFFER | : | 50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.095 % NaN ₃ as a preservative. |
| SHELF LIFE AND STORAGE | : | 18 months at 2-8 °C |
| ANALYTE DESCRIPTION | : | Thyroglobulin is a 660 kDa dimeric protein produced by and used entirely within the thyroid gland to produce the thyroid hormones thyroxine (T4) and triiodothyronine (T3). Thyroglobulin levels in the blood can be used as a tumor marker for certain kinds of thyroid cancer. Thyroglobulin levels in the blood can also be elevated in cases of Graves' disease. |

PARAMETERS TESTED FROM EACH LOT

| | | |
|------------------------------|---|--|
| PRODUCT APPEARANCE | : | Clear liquid, may turn opaque during storage |
| PRODUCT CONCENTRATION | : | 5.0 mg/ml (+/- 10 %) |
| PRODUCT ACTIVITY | : | 80 – 120 % compared to reference in an IFMA-test |
| IEF-RANGE | : | 7.8 – 9.1 |
| PURITY | : | ≥ 95 % |

PARAMETERS DETERMINED ONLY DURING PRODUCT R&D PHASE

| | | |
|------------------------------|---|--|
| CLASS AND SUBCLASS | : | IgG ₁ |
| ASSOCIATION CONSTANT | : | 6.2 x 10 ⁵ 1/Ms |
| DISSOCIATION CONSTANT | : | 1.3 x 10 ⁻⁵ 1/s |
| AFFINITY CONSTANT | : | KA = 4.8 x 10 ¹⁰ 1/M; KD = 2.1 x 10 ⁻¹¹ M (=0.02 nM) |
| DETERMINATION METHOD | : | SPR analysis (BioRad ProteOn XPR36) |
| ANTIGEN | : | Thyroglobulin (Fitzgerald cat.no. 30-AT01, lot#A11102002) |
| CROSS-REACTIVITIES | : | No cross-reactivity against Thyroid Peroxidase, others not tested. |

EPITOPE : N/D

EPITOPE GROUP : N/A

Two antibodies binding to the same, or closely located epitopes, belong to the same group and hence cannot be used as a pair in a sandwich assay. Epitope group numbering does not give any detailed information where the epitope is located.

| PAIR RECOMMENDATIONS | SOLID | LABEL |
|-----------------------------|--------------|------------------|
| | 2805 | 2802, 2803, 2804 |
| | 2803 | 2804, 2805 |

Please note that pair recommendations are based on results obtained in our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations should be taken only as a directive.

| PRODUCT STABILITY | TEMPERATURE, DAYS | RESULT |
|---|--------------------------|---------------|
| Please note that the shelf life given on page one is based on real time stability testing at +2-8 °C in the product buffer. | -70 °C, 21 days | OK |
| | -20 °C, 21 days | OK |
| | +4 °C, 21 days | OK |
| | +25 °C, 21 days | OK |
| | +35 °C, 21 days | OK |
| | +45 °C, 7 days | OK |

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. The maximum duration of the test is 21 days, except for the +45 °C only 7 days.

| pH, 14 DAYS, +4 °C | RESULT |
|---------------------------|---------------|
| 5.0 | OK |
| 6.0 | OK |
| 7.0 | OK |
| 8.0 | OK |

Stability testing is performed to see whether pH affects the antigen binding, charge or composition of the antibody during 14 days at +4 °C.

MISCELLANEOUS : T3 or T4 has no effect on the binding of this antibody.

REFERENCES :

Legal disclaimer

MedixMAB monoclonal antibodies meet their specifications if transported, stored and used according to the instructions for use. See also separate *Notes on use and storage of monoclonal antibodies*. If not otherwise agreed in writing all products are sold under *General Delivery Terms*. Other terms may apply if MedixMAB monoclonal antibodies are purchased from an official distributor of Medix Biochemica products. MedixMAB is a registered trademark of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission. Further information on products and methods is available from medixMAB website and extranet.