

Recombinant antigen EBNA1 for Epstein-Barr Virus (EBV)

CATALOG NUMBER: RAG0007

LOT NUMBER: #

RECOMBINANT ANTIGEN: EBV antigen EBNA1 (Cheng *et al.*, 1991).

DESCRIPTION: the latent nuclear antigen EBNA1 has been prepared as a recombinant antigen fused to a his-tag in its N-terminal. It is produced from the ORF of the BKRF1 gene of EBV.

PRESENTATION: liquid protein solution

SOURCE: *Escherichia coli*

MOLECULAR WEIGHT: determined by SDS-PAGE, the protein band is between molecular markers of 66,200-116,000 Da, while relative molecular mass calculated from amino acid sequence is 63,060.5 Da.

BATCH COMPOSITION:

COMPONENTS	COMPOSITION
his-EBNA1	recombinant antigen with a his-tag in its N-terminus
Storage buffer	20 mM phosphate buffer pH 8, 1 M NaCl and 5 mM EDTA

QUALITY CONTROL:

1. PROTEIN CONCENTRATION DETERMINED ESPECTROPHOTOMETRICALLY

DO₂₈₀ = 0.493
 $A_{0.1\%} (=1 \text{ g/l}) = 0.409$
 CONCENTRATION*: 1.20 mg/ml

* The measurement of the protein concentration has been performed with the theoretical extinction coefficient of the recombinant protein obtained from Gill and von Hippel, 1989

2. PURITY CONTROL IN SDS-PAGE: 12.5%




Figure 1. SDS-PAGE analysis (12.5%) of 5 μ l of recombinant EBNA1. Purity is > 95% as determined by gel electrophoresis. Band which slightly appears above 45 kDa and approx. 28 kDa correspond to this same protein as it is showed in a western blot performed with a his-tag monoclonal antibody.

3. TITRATION CURVE BY AN ELISA ASSAY

The titer has been suggested in reference to an "in-house" ELISA kit performed at Rekom Biotech.

Each end user should carry out his own titration for his particular application.




Figure 2. In this plot, the optical density at 450/620 nm for positive (blue) and negative (gray) IgG sera are compared for each concentration of the recombinant antigen. An appropriate statistical test of significance for the comparison of means between both groups, the Welch's test, is employed. Eligible concentrations for the use of the antigen should present statistically significant differences between positive and negative sera. This happens when the intervals at the top do not overlap and, equivalently, when the p-value at the bottom is below 0.05. In the present figure, all p-values are below 0.05 and thus the intervals do not overlap. Therefore, any of the showed concentrations can be used to distinguish between positive and negative sera.

4. ABSENCE OF PRECIPITATION AFTER A FREEZING AND THAWING CYCLE: ok

LOT SPECIFICATIONS:

- 1. CONCENTRATION:** 1.20 mg/ml
- 2. TOTAL QUANTITY PER ALIQUOT:** 1 mg
- 3. TOTAL VOLUME PER ALIQUOT:** 0.87 ml
- 4. SUGGESTED TITER BY ELISA:** up to 1:20,000, which corresponds to 0.057 $\mu\text{g/ml}$ of protein concentration in plates, for IgG detection.
- 5. STORAGE:** protein is shipped with dry ice. Upon arrival, it should be aliquoted in order to avoid repeated freezing and thawing cycles and stored at -20°C to -80°C
- 6. APPLICATIONS:** ELISA and Western blot assays. Where this product has not been tested for use in a particular technique, this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates.
- 7. OBSERVATIONS:** in some cases, purified proteins run at a molecular weight which is slightly different to the

theoretically calculated molecular weight maybe due to the his-tag presence, which can produce a delay in SDS-PAGE. Proteins should be maintained frozen at high concentrations. The dilution to be performed for ELISA assays should be made with a small quantity of protein, the same day of the experiment. In order to defrost the protein, maintain the aliquot at 25°C without shaking to avoid aggregation. Prior making test dilutions and after defrost the protein, is recommended to remove possible protein aggregates by centrifuging the stock solution, avoiding alterations in the immobilization of the biomolecule to the solid surface.

RELATED PRODUCTS:

p138, p18, p23, p54, ZEBRA.

BIBLIOGRAPHY:

Cheng, H., Y. Foong, C. Sam, U. Prasad and J. Dillner. Epstein-Barr Virus Nuclear Antigen 1 Linear Epitopes That Are Reactive with Immunoglobulin A (IgA) or IgG in Sera from Nasopharyngeal Carcinoma Patients or from Healthy Donors. 1991. *J. Clin. Microbiol.* 29:2180-218.

Gill SC, von Hippel PH. Calculation of protein extinction coefficients from amino acid sequence data. *Anal Biochem.* 1989 Nov 1;182(2):319-26.

Important Notes: During shipment, small volumes of product will occasionally become entrapped in the seal of the product vial. For products with volumes of 200 µl or less, we recommend gently tapping the vial on a hard surface or briefly centrifuging the vial in a tabletop centrifuge to dislodge any liquid in the containers cap.

Although recombinant antigens are expressed in non-pathogenic *E. coli* and bacterial integrity is destroyed during purification, the antigen preparation should be handled as potentially infectious.

NOT FOR DIAGNOSTIC USE, FOR RESEARCH USE ONLY